



CERTIFICATE OF ANALYSIS

Trypsin: EDTA Solution **0.05% Trypsin-0.53mM EDTA** **(Catalog Number 400-150)**

Lot Number: F122018

Date of Manufacture: May2022

Storage Temperature: $\leq -10^{\circ}\text{C}$

Product Expiry: May2024

For Cell culture, Further Manufacturing or Research use Only. Not for Direct Therapeutic Use.

Product description: *Trypsin: EDTA Solution is sterile-filtered through a 0.2 μm filter prior to freezing.*

Test	Methodology	Specification	Analysis
Microbiological Testing			
Sterility	USP<71>, EP 2.6.1		
Bacteria		No Growth	No Growth
Fungi		No Growth	No Growth
Physical Testing			
Osmolality	USP<785>, EP 2.2.35	Report	278 mOsm/kg
pH	USP<791>	Report	7.57

The testing that has been performed as part of this lot release has been reviewed by Quality Assurance personnel and has confirmed that the testing meets the specifications presented on this Certificate of Analysis.

VANESSA YOUNG
Name

QA ASSOCIATE II
Title

03JUN2022
Date