



CERTIFICATE OF ANALYSIS

Human Serum AB

(Off the Clot)

(Catalog Number 100-318)

Lot Number: H121061

Date of Manufacture: Dec2021

Product Expiry: Dec2026

Origin: United States

Storage Temperature: ≤ -10°C

For Cell culture, Further Manufacturing or Research use Only. Not for Direct Therapeutic Use.

Product description: Human Serum AB is collected from healthy male donors of the AB serotype at FDA-licensed facilities located in the United States. Donor units are tested for infectious disease markers prior to processing and found to be non-reactive. Human Serum AB (Off the Clot) is sterile-filtered through a 0.1 µm filter prior to freezing.

Test	Methodology	Specification	Analysis
Biological Testing			
Endotoxin	USP<85>, EP 2.6.14	< 10.0EU/ mL	< 1.68 EU/mL
Hemoglobin	Fleming, AF and Woolf, AJ (1985)	<20.0 mg/ dL	2.5 mg/dL
Microbiological Testing			
Sterility	USP<71>, EP 2.6.1		
Bacteria		No Growth	No Growth
Fungi		No Growth	No Growth
Mycoplasma	USP<63>	Not Detected	Not Detected
Viral Testing			
HBsAg	21CFR 610.40 ABBOTT ChLIA	Non-Reactive	Non-Reactive
Anti-HCV	ABBOTT ChLIA	Non-Reactive	Non-Reactive
Anti-HIV-1/ HIV-2	ABBOTT ChLIA	Non-Reactive	Non-Reactive
Syphilis	ASI RPR	Negative	Negative
HBV-NAT	ROCHE NAT/ PCR	Not Detected	Not Detected
HIV-NAT	ROCHE NAT/ PCR	Not Detected	Not Detected
HCV-NAT	ROCHE NAT/ PCR	Not Detected	Not Detected
Zika Virus RNA	ROCHE NAT/ COBAS	Not Detected	Not Detected
West Nile Virus NAT		Not Detected	Not Detected
Chagas		Not Detected	Not Detected
Physical Testing			
Osmolality	USP<785>, EP 2.2.35	260 – 350 mOsm/ kg	291 mOsm/Kg
pH	USP<791>	Test and Report	7.87



CERTIFICATE OF ANALYSIS

Human Serum AB

(Off the Clot)

(Catalog Number 100-318)

Lot Number: H121061

Test	Methodology	Specification	Analysis
Biochemistry			
Albumin		Test and Report	4.2 g/dL
ALT (SGPT)		Test and Report	9 U/L
AST (SGOT)		Test and Report	13 U/L
Bilirubin, Total		Test and Report	0.2 mg/dL
BUN		Test and Report	14 mg/dL
Calcium		Test and Report	9.2 mg/dL
Chloride		Test and Report	104 mmol/L
Cholesterol		Test and Report	152 mg/dL
Creatinine		Test and Report	1.00 mg/dL
Glucose		Test and Report	100 mg/dL
Phosphorus		Test and Report	3.4 mg/dL
Potassium		Test and Report	4.2 mmol/L
Protein, Total	Biuret	Test and Report	7.0 g/dL
Sodium		Test and Report	139 mmol/L
Triglycerides		Test and Report	75 mg/dL
Uric Acid		Test and Report	5.3 mg/dL

All blood products are collected from stringently screened male donors at FDA-licensed collection centers located in the United States. This product is a pool of > 100 donor units and the viral testing is performed on the individual donor units. All other testing is performed on the final product pool prior to release. Universal precautions should be used when handling this material. Precipitates may develop in this product upon freezing and/or thawing; this occurrence does not impact culture performance. Precipitates may develop in this product upon freezing and/or thawing; this occurrence does not impact culture performance.

The testing that has been performed as part of this lot release has been reviewed by Quality Assurance personnel and has confirmed that the testing meets the specifications presented on this Certificate of Analysis.

 Megha Khatasi
 Name

 QC Supervisor
 Title

 13 JAN 2022
 Date