



CERTIFICATE OF ANALYSIS
BenchMark™ Fetal Bovine Serum
(Catalog Number 100-106) Gamma Irradiated***

Lot Number: A55H74K

Date of Manufacture: Aug2019

Product Expiry: Aug2024

Origin: Honduras

Storage Temperature: ≤ -10°C

For Cell culture, Further Manufacturing or Research use Only. Not for Direct Therapeutic Use.

Product description: BenchMark™ Fetal Bovine Serum is sterile-filtered through a 0.1 µm filter prior to freezing.

Test	Methodology	Specification	Analysis
Biological Testing			
Endotoxin	USP<85>, EP 2.6.14	<10.0 EU/mL	2.39 EU/mL
Hemoglobin	Fleming, AF and Woolf, AJ (1965)	<20.0 mg/dL	15.83 mg/dL
Microbiological Testing			
Sterility	USP<71>, EP 2.6.1		
Bacteria		No Growth	No Growth
Fungi		No Growth	No Growth
Mycoplasma	Barile, MF and Kern, J (1971), EP 2.6.7	Not Detected	Not Detected
Viral Testing			
Cytopathic Agents	9CFR 113.53c[113.46, 113.47]	Not Detected	Not Detected
Hemadsorbing Agents		Not Detected	Not Detected
Bovine Viral Diarrhea Ab Titer*			
Type I		≤1:4	Undiluted
Type II		≤1:4	<Undiluted
Extraneous Viruses (FAb)			
Bluetongue Virus		Not Detected	Not Detected
Bovine Adenovirus		Not Detected	Not Detected
Bovine Parvovirus		Not Detected	Not Detected
Bovine Viral Diarrhea(cytopathic)		Not Detected	Not Detected
Bovine Viral Diarrhea(non-cytopathic)		Test and Report	Not Detected
Bovine Syncytial Virus		Not Detected	Not Detected
Bovine Rabie Virus		Not Detected	Not Detected
Reovirus		Not Detected	Not Detected



CERTIFICATE OF ANALYSIS

BenchMark™ Fetal Bovine Serum

(Catalog Number 100-106) Gamma Irradiated***

Lot Number: A55H74K

Test	Methodology	Specification	Analysis
Physical Testing			
Electrophoretic Pattern	Cellulose Acetate	Normal	Normal
Osmolality	USP<785>, EP 2.2.35	260 – 350 mOsm/ kg	306 mOsm/Kg
pH@RT	USP<791>	6.5 – 8.5	7.16
Biochemistry			
Albumin		Test and Report	1.75 g/dL
Alkaline Phosphatase		Test and Report	390 IU/L
ALT (SGPT)		Test and Report	8 IU/L
AST (SGOT)		Test and Report	56 IU/L
GGT		<10 U/L	7 IU/L
Bilirubin, Total		Test and Report	0.15 mg/dL
BUN		Test and Report	17 mg/dL
Calcium		Test and Report	14.2 mg/dL
Chloride		Test and Report	100 mEq/L
Cholesterol		Test and Report	34 mg/dL
Cortisol		Test and Report	0.330 ug/dL
Creatinine		Test and Report	2.9 mg/dL
Estradiol		Test and Report	16.77 pg/mL
Glucose		Test and Report	88 mg/dL
IgG	ELISA	<0.300 mg/mL**	0.153 mg/mL
Insulin		Test and Report	7.31 uIU/mL
Iron, Serum		Test and Report	157 µg/dL
Phosphorus		Test and Report	9.4 mg/dL
Potassium		Test and Report	11.6 mEq/L
Progesterone		Test and Report	<0.06 ng/mL
Protein, Total	Biuret	2.5 – 6.0 g/dL	3.66 g/dL
Sodium		Test and Report	135 mEq/L
Testosterone		Test and Report	<0.01 ng/mL
T3		Test and Report	110 ng/dL
T4		Test and Report	14.0 ug/dL
Triglycerides		Test and Report	67 mg/dL
Tetracycline		<4 ng/mL	Pass
Uric Acid		Test and Report	2.4 mg/dL



CERTIFICATE OF ANALYSIS
BenchMark™ Fetal Bovine Serum
(Catalog Number 100-106) Gamma Irradiated***

Lot Number: A55H74K

Our fetal bovine serum meets all USDA requirements and has passed ante and post-mortem inspection by a licensed veterinarian. It is collected in countries recognized as being free of foot-and-mouth disease and rinderpest. The origin of all fetal bovine serum is traceable by lot number, date and country of collection. Precipitates may develop in this product upon freezing and/or thawing; this occurrence does not impact culture performance. * Testing not required if Bovine Viral Diarrhea(cytopathic) and Bovine Viral Diarrhea (non – cytopathic) is not detected. **Some geographic regions may test higher than 0.300 mg/mL. If this be the case, refer to GGT results less than 10 U/L to ensure FBS purity. *** Results shown were obtained by carefully performed methods believed to be reliable prior to Gamma Irradiation.

The testing that has been performed as part of this lot release has been reviewed by Quality Assurance personnel and has confirmed that the testing meets the specifications presented on this Certificate of Analysis.

Megha Khayasi

Name

QC Supervisor

Title

04 JUN 2021

Date