



Product: Donor Donkey Serum
Sterile-filtered

Catalog #: 100-151

Lot #: E84H07K

Origin: United States (100%)

Manufacture Date: Jun 2019

Expiration Date: Jun 2022

TEST	METHODOLOGY	SPECIFICATION	RESULTS
Microbiological Testing			
Sterility	USP<71>		
Bacteria		Not Detected	Not Detected
Fungi		Not Detected	Not Detected
Physical Testing			
Osmolality	USP<785>	Report	286 mOsm/kg
pH @ RT	USP<791>	Report	7.9
Biochemical Testing			
Hemoglobin			3.1 mg/dL
Protein, Total			6.0 g/dL

For Cell Culture, further Manufacturing and Research use only. Not for direct Therapeutic use.

[Signature] 14 OCT 2019

Created by/ date

[Signature] 14 OCT 2019

Verified by/ date