



## CERTIFICATE OF ANALYSIS

### Amphotericin B Solution

(Catalog Number 400-104)

Lot Number:

F021022

Date of Manufacture: APR 2021

Product Expiry: APR 2023

Origin: United States

Storage Temperature:  $\leq -10^{\circ}\text{C}$

For Cell Culture, Further Manufacturing or Research use Only. Not for Direct Therapeutic Use.

Product description: Amphotericin B Solution is sterile-filtered through a 0.2  $\mu\text{m}$  filter prior to freezing.

Test	Methodology	Specification	Analysis
<b>Microbiological Testing</b>			
Sterility	USP<71>, EP 2.6.1		
Bacteria		No Growth	No Growth
Fungi		No Growth	No Growth
<b>Physical Testing</b>			
Osmolality	USP<785>, EP 2.2.35	Test and Report	30 mOsm/kg
pH	USP<791>	Test and Report	12.46

The testing that has been performed as part of this lot release has been reviewed by Quality Assurance personnel and has confirmed that the testing meets the specifications presented on this Certificate of Analysis.

Shinica Z Khan

Name

QC Associate

Title

22 APR 2021

Date