

Certificate of Analysis

Created April 17, 2018

Product:

Trypsin: EDTA Solution

0.05% trypsin—0.53mM EDTA

Sterile-filtered

Catalog #: 400-150 Lot #: F04S00J Manufacture Date: Apr 2018 Expiration Date: Apr 2019

TEST	M ETHODOLOGY	SPECIFICATION	RESULTS	
Microbiological Testing	NO. 1			
Sterility	USP<71>	Not Detected	Not Detected	
Physical Testing				
Osmolality	USP<785>	Report	282 m0sm/kg	
pH @ RT	USP<791>	Report	7.64	

Created by

Verified by