

## **Certificate of Analysis**

Created December 08, 2021

Product: Trypsin: EDTA Solution

0.25% -0.53mM Sterile-filtered

Catalog #: 400-151 Manufacture Date: Nov 2021 Lot #: F121075 Expiration Date: Nov 2023

		•		
TEST	<b>M</b> ETHODOLOGY	SPECIFICATION	RESULTS	
Microbiological Testing				
Sterility	USP<71>			
Bacteria		No Growth	No Growth	
Fungi		No Growth	No Growth	
Physical Testing				
Osmolality	USP<785>	Report	284 mOsm/kg	
pH @ RT	USP<791>	Report	7.59	

For Cell Culture, further Manufacturing and Research use only. Not for direct Therapeutic use.

Created by/ date

Verified by/ date