

Certificate of Analysis

Created February 5, 2020

Product:

ITS Media

Sterile-filtered

Catalog #: 400-145 Lot #: F53T00L Manufacture Date: Jan 2020 Expiration Date: Jan 2022

Test	M ETHODOLOGY	SPECIFICATION	RESULTS
Biological Testing			
Endotoxin	USP<85>	Report	<0.5 EU/mL
Microbiological Testing			
Sterility	USP<71>		
Bacteria		Not Detected	Not Detected
Fungi		Not Detected	Not Detected
Physical Testing			
Osmolality	USP<785>	Report	280 mOsm/kg
pH @ RT	USP<791>	Report	8.15

For Cell Culture, further Manufacturing and Research use only. Not for direct Therapeutic use.

Created by/ date

Verified by/ date

05 Feb 2020