

Certificate of Analysis

Created October 07, 2020

Product:

Trypsin: EDTA Solution

0.25% trypsin-0.53mM EDTA

Sterile-filtered

Catalog #: 400-151 Lot #: F85T00L Manufacture Date: SEP 2020 Expiration Date: SEP 2021

Test	M ETHODOLOGY	SPECIFICATION	RESULTS
Microbiological Testing			
Sterility	USP<71>		
Bacteria		Not Detected	Not Detected
Fungi		Not Detected	Not Detected
Physical Testing			
Osmolality	USP<785>	Report	291 m0sm/kg
pH @ RT	USP<791>	Report	7.51

For Cell Culture, further Manufacturing and Research use only. Not for direct Therapeutic use.

DBK 107 OCT 2020

Created by/ date

Werified by/ date