



CERTIFICATE OF ANALYSIS

Human apo-Transferrin

Powder

(Catalog Number 800-130P)

Lot Number: H423014

Date of Manufacture: Jun2023

Product Expiry: Apr2026

Origin: United States

Storage Temperature: 2-8°C

For Cell Culture, further Manufacturing or Research use only. Not for Direct Therapeutic use.

Test	Methodology	Specification	Analysis
Biological Testing			
Bioburden		<10 cfu/mL	Pass
Endotoxin	LAL	<1 EU/mg	<0.05 EU/mg
Microbiological Testing			
Mycoplasma		Not Detected	Not Detected
Viral Testing			
HBsAg		Non-Reactive	Non-Reactive
Anti-HCV		Non-Reactive	Non-Reactive
Anti-HIV-1/ HIV-2		Non-Reactive	Non-Reactive
ALT		Negative	Negative
Syphilis		Negative	Negative
HBV-NAT		Not Detected	Not Detected
HIV-NAT		Not Detected	Not Detected
HAV-NAT		Not Detected	Not Detected
HCV-NAT		Not Detected	Not Detected
Parvo B19-NAT		Non-Reactive	Non-Reactive
Physical Testing			
Appearance		Report	White Crystalline Powder
pH @ RT, 3% Solution		6.5-8.0	6.94
Moisture		≤5.0%	0.49%
Solubility		Readily Soluble at 1% in De-ionized Water	Pass
Optical Density		Report	11.13
Biochemical Testing			
Purity	Cellulose Acetate Electrophoresis	≥98% of total protein	Pass
Protein Analysis		>98%	101.7%
SDS-Polyacrylamide Gel		Conforms	Pass
Iron Estimated ICP		<50 ppm	14 ppm



CERTIFICATE OF ANALYSIS

Human apo-Transferrin

Powder

(Catalog Number 800-130P)

Lot Number: H423014

Results shown were obtained by carefully performed methods believed to be reliable. However, since some results may vary for specific tests depending upon methodology and other variables, it is suggested that tests for which results are particularly important be repeated by the user of this product.

All human blood products are collected from screened donors at FDA-licensed collection centers in the United States.

Single homogenous batch, heat treated at $62^{\circ}\pm 2^{\circ}\text{C}$ for 10 hours and lyophilized from approximately 0.02M NH_4HCO_3 . May contain traces of buffer salts. Dispensed by dry weight.

Store at $2-8^{\circ}\text{C}$. Allow temperature to equilibrate to room temperature prior to use.

The testing that has been performed as part of this lot release has been reviewed by Quality Assurance personnel and has confirmed that the testing meets the specifications presented on this Certificate of Analysis.

VANESSA YOUNG
Name

QA ASSOCIATE II
Title

11 JUL 2023
Date