



CERTIFICATE OF ANALYSIS

Viral Transport Medium

(Catalog Number 600-102)

Parent Lot Number: M021004

Date of Manufacture: Apr2021

Sub-Lot Number: M021004-1

Product Expiry: Apr2022

For Transport of Specimens only.
Not to be taken internally.

For In Vitro Diagnostic Use
Storage Temperature: 2°C to 8°C DO NOT FREEZE

Product description: *Viral Transport Medium is sterile-filtered through a 0.2 µm filter prior to refrigeration.*

| Test | Methodology | Specification | Analysis |
|--------------------------------|-------------|---------------|--------------|
| Biological Testing | | | |
| Endotoxin | USP <85> | Report | <0.050 EU/mL |
| Microbiological Testing | | | |
| 37°C Sterility Check | | No Growth | No Growth |
| Physical Testing | | | |
| pH@RT | USP<791> | Report | 7.64 |
| Osmolality | USP<785> | Report | 284 |

The testing that has been performed as part of this lot release has been reviewed by Quality Assurance personnel and has confirmed that the testing meets the specifications presented on this Certificate of Analysis.

Chuntug Francisco

QC Associate

Name 13 APR 2021

Title

Date