

Product: **Vivify -1X™ CryoStorage Medium**
 0.2 µm Sterile-filtered

Catalog #: 600-100
Lot #: M35S00K

Manufacture Date: Oct 2019
Expiration Date: Oct 2021

TEST	METHODOLOGY	SPECIFICATION	RESULTS
Microbiological Testing			
Sterility	USP<71>		
Bacteria		Not Detected	Not Detected
Fungi		Not Detected	Not Detected
Physical Testing			
Osmolality	USP<785>	Report	498 mOsm/kg
pH @ RT	USP<791>	Report	7.47

For Cell Culture, further Manufacturing and Research use only. Not for direct Therapeutic use.

[Signature] 08 NOV 2019

Created by/ date

[Signature] 08 NOV 2019

Verified by/ date