



CERTIFICATE OF ANALYSIS

Viral Transport Medium

(Catalog Number 600-102)

Parent Lot Number: M61S00L

Date of Manufacture: May2020

Sub-Lot Number: M61S00L-1, M61S00L-2, M61S00L-3, M61S00L-4, M61S00L-5, M61S00L-6, M61S00L-7, M61S00L-8, M61S00L-9, M61S00L-10, M61S00L-11, M61S00L-12, M61S00L-13, M61S00L-14, M61S00L-15, M61S00L-16

Product Expiry: May2021

For Transport of Specimens only.
Not to be taken internally.

For In Vitro Diagnostic Use
Storage Temperature: 2°C to 8°C DO NOT FREEZE

Product description: *Viral Transport Medium is sterile-filtered through a 0.2 µm filter prior to refrigeration.*

Test	Methodology	Specification	Analysis
Biological Testing			
Endotoxin	USP <85>	Report	<0.500 EU/mL
Microbiological Testing			
37°C Sterility Check		No Growth	No Growth
Physical Testing			
pH@RT	USP<791>	Report	7.67
Osmolality	USP<785>	Report	275 mOsm/Kg

This Device has not been cleared or approved by the FDA.

The testing that has been performed as part of this lot release has been reviewed by Quality Assurance personnel and has confirmed that the testing meets the specifications presented on this Certificate of Analysis.

Megha Khawari
Name
03 Jun 2020
Date

QA Associate III
Title