



## CERTIFICATE OF ANALYSIS

### Viral Transport Medium

(Catalog Number 600-102)

Parent Lot Number: M75S00L

Date of Manufacture: Jul2020

Sub-Lot Number: M75S00L-16

Product Expiry: Jul2021

For Transport of Specimens only.  
Not to be taken internally.

For In Vitro Diagnostic Use  
Storage Temperature: 2°C to 8°C DO NOT FREEZE

Product description: *Viral Transport Medium is sterile-filtered through a 0.2 µm filter prior to refrigeration.*

Test	Methodology	Specification	Analysis
<b>Biological Testing</b>			
Endotoxin	USP <85>	Report	<0.540 EU/mL
<b>Microbiological Testing</b>			
37°C Sterility Check		No Growth	
<b>Physical Testing</b>			
pH@RT	USP<791>	Report	7.40
Osmolality	USP<785>	Report	280 mOsm/Kg

The testing that has been performed as part of this lot release has been reviewed by Quality Assurance personnel and has confirmed that the testing meets the specifications presented on this Certificate of Analysis.

Megha Kharasi

Name

QA Associate III

Title

27 Jul 2020

Date