



CERTIFICATE OF ANALYSIS

Saline Solution

(Catalog Number 600-107)

Parent Lot Number: M79S00L

Date of Manufacture: Aug2020

Sub-Lot Number: M79S00L-7,M79S00L-8,M79S00L-9,M79S00L-10,M79S00L-11,M79S00L-12,M79S00L-13,M79S00L-14,M79S00L-15,M79S00L-16, M79S00L-17,M79S00L-18,M79S00L-19,M79S00L-20

Product Expiry: Aug2022

For Transport of Specimens only.
Not to be taken internally.

For In Vitro Diagnostic Use
Storage Temperature: Store at 15 - 30°C

Product description: Saline Solution is sterile-filtered through a 0.2 µm filter prior to storage.

| Test | Methodology | Specification | Analysis |
|--------------------------------|-------------|---------------|--------------|
| Biological Testing | | | |
| Endotoxin | USP <85> | Report | <0.500 EU/mL |
| Microbiological Testing | | | |
| 37°C Sterility Check | | No Growth | No Growth |
| Physical Testing | | | |
| pH@RT | USP<791> | Report | 6.34 |
| Osmolality | USP<785> | Report | 270 mOsm/Kg |

The testing that has been performed as part of this lot release has been reviewed by Quality Assurance personnel and has confirmed that the testing meets the specifications presented on this Certificate of Analysis.

Megha Khawari

Name

QA Associate III

Title

26 Aug 2020

Date