



**CERTIFICATE OF ANALYSIS**  
**Saline Solution**  
**(Catalog Number 600-107)**

**Parent Lot Number:** M87S00L

**Date of Manufacture:** Aug2020

**Sub-Lot Number:** M87S00L-1, M87S00L-2, M87S00L-3, M87S00L-4, M87S00L-5, M87S00L-6, M87S00L-7, M87S00L-8, M87S00L-9, M87S00L-10, M87S00L-11, M87S00L-12, M87S00L-14, M87S00L-15, M87S00L-16, M87S00L-17, M87S00L-18, M87S00L-19, M87S00L-20, M87S00L-21, M87S00L-22, M87S00L-23, M87S00L-24, M87S00L-25

**Product Expiry:** Aug2022

For Transport of Specimens only.  
Not to be taken internally.

For In Vitro Diagnostic Use  
**Storage Temperature:** Store at 15 - 30°C

**Product description:** Saline Solution is sterile-filtered through a 0.2 µm filter prior to storage.

Test	Methodology	Specification	Analysis
<b>Biological Testing</b>			
Endotoxin	USP <85>	Report	<0.524 EU/mL
<b>Microbiological Testing</b>			
37°C Sterility Check		No Growth	No Growth
<b>Physical Testing</b>			
pH@RT	USP<791>	Report	5.88
Osmolality	USP<785>	Report	270 mOsm/kg

The testing that has been performed as part of this lot release has been reviewed by Quality Assurance personnel and has confirmed that the testing meets the specifications presented on this Certificate of Analysis.

Megha Khalasi  
Name  
16 SEP 2020  
Date

QA Associate III  
Title