



**CERTIFICATE OF ANALYSIS**  
**Saline Solution**  
**(Catalog Number 600-107)**

**Parent Lot Number:** M91S00L

**Date of Manufacture:** Sep2020

**Sub-Lot Number:** M91S00L-1, M91S00L-2, M91S00L-3, M91S00L-4, M91S00L-5, M91S00L-6, M91S00L-7, M91S00L-8, M91S00L-9, M91S00L-10, M91S00L-11, M91S00L-12

**Product Expiry:** Sep2022

For Transport of Specimens only.  
Not to be taken internally.

For In Vitro Diagnostic Use  
**Storage Temperature:** Store at 15 - 30°C

**Product description:** Saline Solution is sterile-filtered through a 0.2 µm filter prior to storage.

Test	Methodology	Specification	Analysis
<b>Biological Testing</b>			
Endotoxin	USP <85>	Report	<0.500 EU/mL
<b>Microbiological Testing</b>			
37°C Sterility Check		No Growth	No Growth
<b>Physical Testing</b>			
pH@RT	USP<791>	Report	6.84
Osmolality	USP<785>	Report	269 mOsm/kg

The testing that has been performed as part of this lot release has been reviewed by Quality Assurance personnel and has confirmed that the testing meets the specifications presented on this Certificate of Analysis.

Megha Khalasi

QA Associate III

Name

Title

02 OCT 2020

Date