



GEMINI
BIO-PRODUCTS

Certificate of Analysis

Created December 16, 2020

Product: **NeuroPlex™ Serum-Free Neuronal Medium**
Sterile-filtered

Catalog #: 600-300
Lot #: M95S00L

Manufacture Date: NOV 2020
Expiration Date: NOV 2022

TEST	METHODOLOGY	SPECIFICATION	RESULTS
Biological Testing			
Endotoxin	USP<85>	Report	<0.592 EU/mL
Microbiological Testing			
Sterility	USP<71>		
Bacteria		No Growth	No Growth
Fungi		No Growth	No Growth
Physical Testing			
Osmolality		240-280 mOsm/kg	244 mOsm/kg
pH		7.25-7.35	7.31

For Cell Culture, further Manufacturing and Research use only. Not for direct Therapeutic use.

DBK 16 DEC 2020

Created by/ date

MK 16 DEC 2020

Verified by/ date