



CERTIFICATE OF ANALYSIS
BenchMark™ Fetal Bovine Serum
(Catalog Number 100-106)

Lot Number: A423013

Date of Manufacture: MAR2023

Product Expiry: MAR2028

Origin: Mexico

Storage Temperature: ≤ -10°C

For Cell Culture, further Manufacturing or Research use only. Not for direct Therapeutic use.

Product Description: BenchMark™ Fetal Bovine Serum is sterile filtered through a 0.1 µm filter prior to freezing.

Test	Methodology	Specification	Analysis
Biological Testing			
Endotoxin	USP <85> & EP 2.6.14	<10.0 EU/mL	<0.1 EU/mL
Hemoglobin	USP <90>	<20.0 mg/dL	9.72 mg/dL
Microbiological Testing			
Sterility	USP <71> & EP 2.6.1		
Bacteria		No Growth	No Growth
Fungi		No Growth	No Growth
Mycoplasma	Barile and Kern & Hoechst Fluorochrome Stain	Not Detected	Not Detected
Viral Testing	9CFR 113.53c [113.46, 113.47]		
Cytopathic Agents		Not Detected	Not Detected
Hemadsorbing Agents		Not Detected	Not Detected
Extraneous Viruses (FAB)			
Bluetongue Virus		Not Detected	Not Detected
Bovine Adenovirus		Not Detected	Not Detected
Bovine Parvovirus		Not Detected	Not Detected
Bovine Viral Diarrhea (Cytopathic)		Not Detected	Not Detected
Bovine Respiratory Syncytial Virus		Not Detected	Not Detected
Rabies Virus		Not Detected	Not Detected
Reovirus		Not Detected	Not Detected
Physical Testing			
Electrophoretic Pattern	Cellulose Acetate Electrophoresis	Normal	Normal
Osmolality	USP <785> & EP 2.2.35	260 – 350 mOsm/ kg	307 mOsm/kg H ₂ O
pH	USP <791> & EP 2.2.3	6.5 – 8.5	7.23
Biochemistry			
Albumin		Test and Report	2.1 g/dL
Alkaline Phosphatase		Test and Report	272 IU/L
ALT (SGPT)		Test and Report	5 IU/L



CERTIFICATE OF ANALYSIS
BenchMark™ Fetal Bovine Serum
(Catalog Number 100-106)

Lot Number: A423013

Test	Methodology	Specification	Analysis
AST (SGOT)	Chemistry Analyzer	Test and Report	25 IU/L
GGT		<10 U/L	5 IU/L
Bilirubin, Total		Test and Report	0.2 mg/dL
BUN		Test and Report	16 mg/dL
Calcium	ELISA	Test and Report	13.1 mg/dL
Chloride		Test and Report	98 mmol/L
Cholesterol		Test and Report	31 mg/dL
Cortisol		Test and Report	0.310 µg/dL
Creatinine		Test and Report	2.9 mg/dL
Glucose		Test and Report	81 mg/dL
IgG		<0.300 mg/mL*	0.176 mg/mL
Insulin		Test and Report	11.33 µIU/mL
Iron, Serum		Test and Report	171 µg/dL
Phosphorus		Test and Report	9.0 mg/dL
Potassium	USP <1057> & EP 2.5.33	Test and Report	11.6 mmol/L
Progesterone		Test and Report	<0.1 ng/mL
Protein, Total		2.5 – 6.0 g/dL	3.5 g/dL
Sodium		Test and Report	136 mmol/L
Testosterone		Test and Report	0.02 ng/mL
T3		Test and Report	107.0 ng/dL
T4		Test and Report	16.0 µg/dL
Triglycerides		Test and Report	64 mg/dL
Uric Acid		Test and Report	2.4 mg/dL

Our fetal bovine serum meets all USDA requirements and has passed ante and post-mortem inspection by a licensed veterinarian. It is collected in countries recognized as being free of foot-and-mouth disease and rinderpest. The origin of all fetal bovine serum is traceable by lot number, date, and country of collection. Precipitates may develop in this product upon freezing and/or thawing; this occurrence does not impact culture performance. *Some geographic regions may test higher than 0.300 mg/mL. If this be the case, refer to GGT results less than 10 U/L to ensure FBS purity.

The testing that has been performed as part of this lot release has been reviewed by Quality Assurance personnel and has confirmed that the testing meets the specifications presented on this Certificate of Analysis.

Jaqueline Wilson
Name
QA Specialist
Title

27 JUN 2023
Date